
Original Operating Instructions




Combustion Calorimeter

CWD2005, CWD2005 Plus



UNION Instruments GmbH

Zeppelinstraße 42
76185 Karlsruhe
Germany

 +49 (0)721-680381-0
 +49 (0)721-680381-33
 support@union-instruments.com
 <http://www.union-instruments.com>

© 2016

This documentation is protected by copyright. All rights, in particular the rights of translation, reprint, extraction of figures, radio transmission, and duplication by photomechanical or other means and of storage in data processing systems, in whole or in part, are reserved.

Technical data subject to change.

Table of contents

Table of contents	3
1 Technical data	5
1.1 Dimensions.....	5
1.2 Device overview	6
1.3 Voltage supply	6
1.4 Interfaces	6
1.5 Display times of combustion value measurement	7
1.6 Gas inputs	7
1.7 Calibration gas / Test gas	7
1.8 Process gas.....	8
1.9 Linearity and measuring ranges.....	8
1.10 Environmental conditions.....	9
2 EU Declaration of Conformity.....	11
3 Safety information	13
3.1 Warning information and symbols.....	13
3.2 Principle, intended use	14
3.3 Personnel and qualification.....	14
3.4 Safety information.....	15
3.4.1 General safety information	15
3.4.2 Information on specific hazards.....	15
3.4.3 Note on explosion protection	16
3.5 Safety precautions on the owner side.....	16
3.5.1 Recurring operator training.....	17
3.5.2 Performing a workplace hazard analysis	17
4 Protective equipment.....	19
4.1 Main switch.....	19
4.2 Safety switch	19
4.3 Thermal cutoff	20
4.4 Solenoid valve	20
4.5 Fan.....	20
4.6 Door enclosure.....	20
4.7 Markings and warning information	21
5 Description and connections.....	23
5.1 Enclosure connections outside	23
5.2 Connections and components inside	25
5.3 Accessories	26
6 Transport, installation, and acceptance	27
6.1 Transport.....	27
6.2 Environmental conditions.....	28
6.2.1 Storage conditions.....	28
6.3 Installing and connecting	29
6.3.1 Installation location.....	29
6.3.2 Room ventilation	30
6.3.3 Wall mounting	31
6.3.4 Process gas	33
6.3.5 Carrier gas supply	34
6.3.6 Calibration gas	35
6.3.7 Flue gas	36

6.3.8	Electrical connection	37
6.3.9	Electrical interfaces	38
6.3.10	Connector assignment Input-Output IOexternal.....	39
6.4	Commissioning after installation	42
6.4.1	Removing/attaching transport securing devices.....	43
6.5	Documentation	44
7	Commissioning/Switching on.....	45
8	Description of the work stations/operator control elements.....	47
9	Operation	49
9.1	Operation of membrane keyboard/Description of display	49
9.2	Basic operation.....	51
9.3	Available displays	51
9.4	General information	52
9.5	Menu structure.....	53
9.6	Main menu	54
9.6.1	Main menu - Options	55
9.6.2	Main menu - Trend	80
9.6.3	Main menu - Event list	85
9.6.4	Main menu - Device information	86
10	Decommissioning / Switching off.....	89
11	Maintenance.....	91
11.1	Preparations	91
11.2	Maintenance work/Inspection	92
12	Troubleshooting	95
12.1	Preparations for troubleshooting	96
12.2	Changing/replacing fuses	97
12.3	Unstable measured value	97
12.4	Drift of measured value.....	97
12.4.1	Incorrect ignition	97
12.4.2	Error/status messages.....	98
13	Service	101
14	Disposal	103
15	Spare parts.....	105
16	Appendix	107
	Data Structure Profibus.....	107
	Data structure serial interface RS 232 (option)	108
	Index	109
	List of figures	110
	List of tables.....	110

1 Technical data

1.1 Dimensions




Fig. 1.1: Housing dimensions

Weight: approx. 54 kg

1.2 Device overview




Fig. 1.2: Markings and warning information

1. Warning information for electricity
2. Name plate

1.3 Voltage supply

Voltage:	230 Volt or 115 Volt 50/60 Hz alternating current
Power consumption:	185 VA max.
Protection class:	I
Degree of protection:	IP50

1.4 Interfaces

Interfaces:	RS232, Profibus, Fieldbus, Industr. Ethernet
Analog outputs:	3, 4 - 20 mA for heating value, Wobbe, and density
Digital outputs:	3 relays

1.5 Display times of combustion value measurement¹

Dead time:	3 seconds
50% time:	7 seconds
90% time:	CWD2005: 20 seconds, CWD2005 PLUS: 15 seconds
99% time:	CWD2005: 60 seconds, CWD2005 PLUS: 55 seconds

1.6 Gas inputs


Process gas:	1, optionally 2 for dual measuring ranges
Calibration inputs:	1, optionally 2 for dual measuring ranges
Test gas inputs:	1
Fast loop:	optional
Gas connections:	Compression fitting 6 mm

1.7 Calibration gas / Test gas

Calibration gas:	Dependent on the process gas
Calibration interval:	Dependent on the process gas
Calibration duration:	Process lasts up to 20 minutes, gas flow approx. 10 minutes
Gas consumption:	Approx. 7 l gas per calibration (dependent on the calibration gas)
Input pressure, max:	40 mbar
Input pressure, min.:	20 mbar, dependent on the gas

¹ The display times are measured from the time the new gas reaches the burner. Display times are based on measurements with pure methane.

1.8 Process gas

	NOTE
	<p>The process gas must be free of condensate, dust and other contaminants, such as naphthalene.</p>



Gas connection inlet pressure: 20 - 40 mbar
 Gas consumption: 30 - 40 l/h (min. rel. density 0.50 with 0.55 mm nozzle)

Measuring range [MJ/m ³] / Accuracy of Wobbe index [± MBE]												
CWD device series	CWD2005		CWD2005 CT		CWD2005 PLUS		CWD2005 DPC		CWD2005 SPC		CWD2000 Ex	
Flare gas	0 – 15	3.0	—	—	—	—	0 – 15	3.0	0 – 15	3.0	—	—
Blast furnace gas	3.5 – 6	3.0	—	—	3.5 – 6	3.0	3.5 – 6	3.0	3.5 – 6	3.0	—	—
Converter gas	4.5 – 9	1.5	—	—	—	—	4.5 – 9	1.5	4.5 – 9	1.5	—	—
Mixed gas	5 – 10	2.0	—	—	5 – 10	2.0	5 – 10	2.0	5 – 10	2.0	—	—
Coke oven gas	15 – 30	1.5	—	—	15 – 30	1.5	15 – 30	1.5	15 – 30	1.5	—	—
Biogas	25 – 35	1.5	—	—	25 – 35	1.5	25 – 35	1.5	25 – 35	1.5	—	—
Natural gas	25 – 48	1.5	30 – 47	1.0	25 – 48	1.0	25 – 48	1.0	25 – 48	1.0	25 – 47	1.0
Refinery gas	25 – 50	1.5	—	—	25 – 50	1.5	25 – 50	1.5	25 – 50	1.5	—	—
Liquefied petroleum gas (LPG)	40 - 90	1.5	—	—	40 - 90	1.5	40 - 90	1.5	40 - 90	1.5	—	—

1.9 Linearity and measuring ranges

The measuring ranges cannot be utilized from 0% to 100%. The range is dependent on the gas composition. Ranges of 45-100 % are typical for a measuring range. Hydrogen content in the gas increases the measuring range span. Inert gases such as N₂, O₂, or CO₂, reduce the measuring range span.

1.10 Environmental conditions

	 NOTICE
	When the combustion calorimeter is used outside of the environmental conditions, additional measures (air conditioning of the combustion calorimeter, etc.) must be agreed on with UNION Instruments GmbH!

Installation location:	Installation room required (☞ <i>Chapter 4</i>)
Room temperature:	5 – 40 °C
Temperature change	≤ 5 °C per hour
Air humidity:	0 – 95 % rel. humidity
Ambient pressure:	800 – 1100 hPa (0.8 – 1.1 bar)

2 EU Declaration of Conformity



Der Hersteller / The manufacturer

UNION Instruments GmbH
Zeppelinstraße 42
76185 Karlsruhe

erklärt hiermit, dass folgend bezeichnete Produkte / hereby declares, that following named products:

Produktbezeichnung:	Verbrennungskalorimeter	Gerätegruppe: CWD2005
Product name:	Calorimeter	device group: CWD2005

konform sind mit den Anforderungen, die in der EU – Richtlinie festgelegt sind / are compliant with the requirements as defined in the EU directive:

2014/30/EU	Elektromagnetische Verträglichkeit
2014/30/EU	Electromagnetic compatibility
2014/35/EU	Elektrische Betriebsmittel zur Verwendung innerhalb bestimmter Spannungsgrenzen
2014/35/EU	electrical equipment designed for use within certain voltage limits

Angewandte harmonisierte Normen / Used harmonized standards:

EN 61010-1:2010	Sicherheitsbestimmungen für elektrische Mess-, Steuer-, Regel- und Laborgeräte - Teil 1: Allgemeine Anforderungen Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements
EN 61326-1:2013	Elektrische Mess-, Steuer-, Regel- und Laborgeräte - EMV-Anforderungen - Teil 1: Allgemeine Anforderungen Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements

Name des Dokumentationsbevollmächtigten:	Schlichter
Name delegate of documentation	

Adresse des Dokumentationsbevollmächtigten:	siehe Adresse des Herstellers
address delegate of documentation	see address of manufacturer

Bei einer nicht autorisierten Änderung des Gerätes verliert diese Erklärung ihre Gültigkeit. / Any unauthorized modification of the device results in invalidity of this declaration.




EU Declaration of Conformity


3 Safety information


3.1 Warning information and symbols

The operating instructions use the following nomenclature and symbols for especially important information:



	 DANGER
	<p>For an immediate danger that can lead to serious physical injury or death!</p>

	 WARNING
	<p>For a potentially dangerous situation that can lead to serious physical injury or death!</p>

	 NOTICE
	<p>For a potentially dangerous situation that can lead to minor physical injury! This may also be used for warnings of property damage!</p>

	NOTE
	<p>For information that can improve the operation of the combustion calorimeter or contribute to prevention of property damage.</p>

3.2 Principle, intended use

	 WARNING
	<p>Intended use also includes observance of these operating instructions! In addition to the following safety information, the safety information of linked system components must also be observed!</p> <p>Additional equipment or accessory parts not installed, supplied, or made by UNION Instruments GmbH require manufacturer's approval by UNION Instruments GmbH! Any warranty is otherwise voided!</p>

The combustion calorimeter is a digitally controlled calorimeter. It measures according to the dry measurement principle. The heat is transferred from the burner to the measuring elements through air.

The combustion calorimeter is a sensor system for determining the combustion value of different process gases. The analysis serves the process control.

The combustion calorimeter is intended for use in weather-proof areas and for fixed mounting and installation.

The combustion calorimeter is not suitable for operation in hazardous areas.

In the case of toxic or explosive gases, the safety provisions applicable at the installation location must be complied with.

Any use beyond this is regarded as intended use. The manufacturer is not liable for damage resulting from this. In this case, the risk is borne solely by the installation engineer/commissioning engineer/owner/operator.



Only skilled persons are permitted to make changes to the combustion calorimeter (mechanical/electrical/pneumatic changes).

3.3 Personnel and qualification

Gas connections and work on the electrical equipment of the combustion calorimeter may only be carried out by a skilled person in compliance with safety provisions, especially for hazardous areas.

3.4 Safety information


3.4.1 General safety information

	 WARNING
	<p>Only operate the combustion calorimeter when all protective equipment is present and operational!</p> <p>Further safety information:</p> <p>☞ <i>Before the corresponding chapters!</i></p>



3.4.2 Information on specific hazards

	 DANGER
	<ul style="list-style-type: none">• After installation, all gas-conveying parts must be checked for leak tightness according to national regulations.• Any type of repair that requires opening of the protective cover may only be carried out by instructed qualified personnel!

3.4.3 Note on explosion protection


	<h2 style="margin: 0;">NOTE</h2>
	<p>The operation/installation of a UNION combustion calorimeter does not produce a hazardous area because combustible gas does not escape from the calorimeter and form an explosive atmosphere.</p> <p>No combustible gas escapes from the UNION combustion calorimeter because:</p> <ul style="list-style-type: none"> • in conformity with the intended use of the combustion calorimeter, the gas is combusted in the device, • the gas supply in the device is shut off as soon as the flame in the burner goes out or if the fan fails, and • the gas-conveying lines can be regarded as technically leak-proof on a continuing basis (compression fittings) or through a systematically recurring testing of tubes and membranes described in the operating instructions. <p>All gas inputs in the combustion calorimeter are provided with solenoid valves. The solenoid valves close automatically in the event of a fault during combustion or ignition of the gas mixture.</p> <p>The solenoid valves are closed in current-free state.</p>

3.5 Safety precautions on the owner side


	 <h2 style="margin: 0;">WARNING</h2>
	<ul style="list-style-type: none"> • The owner must provide suitable protective equipment for the combustion calorimeter that can reliably prevent injuries to personnel, e.g., from escaping gas. • Exhaust discharged process gas to a safe environment! • Tripping hazard from improperly routed supply lines!

Other owner-side safety precautions *☞ corresponding chapter!*

3.5.1 Recurring operator training

	NOTE
	Country-specific regulations regarding recurring training of operators must be observed by the owner, particularly with regard to the handling of hazardous areas, gases, and electrical equipment.




3.5.2 Performing a workplace hazard analysis

	NOTE
	Depending on the national regulations and, if necessary, independent of the CE marking of this combustion calorimeter, the owner must prepare a workplace hazard analysis and specify personal safety equipment for different phases of operation.

Deviations from these operating instructions may occur due to further technical developments. If you desire additional information or if specific problems arise that are not covered in detail in this manual, you will receive information by contacting the following address:

UNION Instruments GmbH

Zeppelinstraße 42
76185 Karlsruhe
Germany

 +49 (0)721-680381-0
 +49 (0)721-680381-33
 support@union-instruments.com
 <http://www.union-instruments.com>



4 Protective equipment

4.1 Main switch




Fig. 4.1: Main switch (example)

4.2 Safety switch

The safety switch disconnects the ignition transformer from the voltage supply when the upper door is opened.



Fig. 4.2: Figure Safety switch (example)

4.3 Thermal cutoff



NOTE

The tripping temperature of the thermal cutoff is 72 °C.

When tripped, the thermal cutoff disconnects the 24 V voltage supply of the solenoid valves.

4.4 Solenoid valve

In the event of malfunctions, the solenoid valves of the combustion calorimeter close automatically. A restart of the combustion calorimeter is necessary.

4.5 Fan

The fan mixes exhaust gases produced during combustion with fresh air and ensures an appropriate volume flow. The fan is speed-monitored. If the fan fails, the combustion calorimeter goes to a safe state.

4.6 Door enclosure

Protective covers of the housing separate electrical components and hot surfaces of the environment.

4.7 Markings and warning information




Fig. 4.3: Markings and warning information (example)

Item No.	Description
1	Warning information: hot surface (internal)
2	Warning information: electricity (external)




Fig. 4.4: Warning information on electrical connection plate (example)



5 Description and connections

5.1 Enclosure connections outside




Fig. 5.1: Housing connections

Item No.	Description
1	Cable glands for power supply
2	Cable glands for signals
3	Main switch
4	Filter cover

Description and connections

Item No.	Description
5	Door interlock (4 pieces)
6	Display
7	Flue exhaust
8	Cover for analog signals (4 pieces)
9	Fast Loop output (optional)
10	Carrier gas input
11	Calibration gas input (SV X14/3-4)
12	Calibration gas input (SV X14/1-2)
13	Process gas input (SV X11/1-2)
14	Cover for interface connections (2 pieces)
15	USB connection

5.2 Connections and components inside







Fig. 5.2: Housing without doors, example

Item No.	Description
1	Thermal cutoff (behind the exhaust gas pipe, on the rear wall)
2	Thermal element
3	Density measuring cell
4	Precision pressure controller
5	Electronic input and output signals
6	Safety switch for ignition transformer
7	Power supply distribution rail, example
8	Fan



5.3 Accessories


 A large yellow warning triangle with a black exclamation mark inside.	 WARNING
	<p>Risk of injury/damage!</p> <p>The use of non-approved accessories may cause damage and endanger persons. Any warranty is voided in this case! The owner is then liable for damage that occurs!</p> <p>Only use genuine accessories or accessories approved by UNION Instruments GmbH.</p>

6 Transport, installation, and acceptance



	NOTE
	<p>The combustion calorimeter is generally commissioned by UNION Instruments GmbH or correspondingly qualified service technicians.</p> <p>When it is not commissioned by UNION Instruments GmbH (e.g., internal transport/resale), the suitable procedure must be agreed with UNION Instruments GmbH (<i>☞ Chapter 13 Service</i>).</p>

6.1 Transport


 	WARNING
	<p>Tipping over or dropping of the combustion calorimeter from the pallet or load carrying means may cause injuries!</p> <p>Use suitable load carrying means when unpacking and transporting.</p> <p>Check slings, if applicable, for adequate load bearing capacity and sound condition and fasten them carefully!</p> <p>Never walk or stand under suspended loads!</p>

	NOTE
	<p>If strong shocks occur during transport, these may damage the housing. Therefore, check the transport container for damage before opening!</p> <p>In case of transport damage that is indicative of improper handling, a damage assessment by the transport carrier (rail, mail, shipping company) must be arranged within seven days.</p> <p>Ensure before starting or restarting transport that all transport securing devices are attached.</p>

6.2 Environmental conditions

	 NOTICE
	<p>Comply with environmental conditions for storage and installation! Contact UNION Instruments GmbH if the combustion calorimeter is stored for longer than 3 months or must be operated or stored outside the prescribed environmental conditions!</p>

6.2.1 Storage conditions

	NOTE
	<ul style="list-style-type: none">• Ensure that the combustion calorimeter is free of gas/moisture residues.• Frozen condensate water in the combustion calorimeter may cause damage.

Ambient temperature: 0 – 60 °C
Air humidity: 0 – 95% relative humidity
Ambient pressure: 700 – 1400 hPa (0.7 – 1.4 bar)



6.3 Installing and connecting


6.3.1 Installation location

The installation location of the combustion calorimeter must meet the following conditions:

- Clean room that may be used only for purposes of gas analysis and measurements.
- At least 50 cm working space on the left side
- Lockable windows
- Protected from direct influence by the weather and direct solar radiation
- Stable room temperature
- Provided with clean and adequate ambient air for unbiased measurement result (combustion calorimeter requires approximately 30 m³/h air)
- Adequate load bearing capacity of the wall is ensured
- Mounting on a solid wall

6.3.2 Room ventilation



	 WARNING
	<p>Risk of injury due to temperature of exhaust gas/housing! The exhaust gas temperature is between 8 - 20 °C above the housing temperature! Use personal safety equipment to prevent burns!</p>


	NOTE
	<p>In the case of unfavorable circulation/installation conditions, appropriate deflecting sheets must be provided that prevent fresh air from flowing directly onto the combustion calorimeter.</p>

The following conditions must be met:

- No direct fresh air introduction in the combustion calorimeter.
- Exhaust gases must be discharged to the environment protected from air draft using a suitable pipe/tube and with the help of corresponding fresh air mixing. Connection points must be checked for leak tightness.

6.3.3 Wall mounting

	 WARNING
	<p>Risk of injury due to the weight of the device! For weight, see technical data! Specify measures to prevent the device from falling down and use suitable hoisting devices!</p>

	NOTE
	<p>Adequate distance must be provided between the side wall of the installation location and the combustion calorimeter for service and maintenance.</p>

The combustion calorimeter is intended to be wall-mounted. The wall brackets are provided and must be secured to the wall.

The wall used for mounting must be sufficiently sturdy to bear the weight of the combustion calorimeter.

Cross-struts are installed in a fixed manner on the back of the combustion calorimeter as the counterpart of the wall brackets.



Fig. 6.1: Wall mounting







Fig. 6.2: Dimensions wall mounting

6.3.4 Process gas

	 <h2 style="margin: 0;">WARNING</h2>
	<p>Gas connections may only be installed by qualified personnel! Exhaust gases must be discharged by the owner to a safe environment!</p>

	<h2 style="margin: 0;">NOTE</h2>
	<ul style="list-style-type: none"> • Connection parts must be clean and free of residues. Contaminants may reach inside the combustion calorimeter and may cause false measurements and/or damage. • Process gas connection / compression fitting on side of combustion calorimeter • Gas input pressure depending on gas between 20 - 40 mbar. ☞ <i>Chapter 1</i> Technical data • The input pressure for the gas connections must not exceed the specifications of the technical data of the combustion calorimeter. • Process gas must be free of contaminants and condensate. • Each connection point must be carefully checked for leak tightness. When leaks are present, the system draws air and indicates false measured values. • Do not use sealant for sealing the gas connections. Sealant ingredients may falsify the measurement result. • Only suitable lines may be used.



6.3.5 Carrier gas supply


NOTE







- Combustion calorimeters can be subsequently converted to a carrier gas supply. Contact the manufacturer regarding this.
- For process gases that do not have stable combustion, a carrier gas can be added.
- Carrier gases maintain the combustion and exert no increased contribution toward the combustion value of the process gas.


6.3.6 Calibration gas

	 <h2 style="margin: 0;">WARNING</h2>
	<p>Gas connections may only be installed by qualified personnel!</p> <p>If a pressure reducer is not installed, escaping calibration gas must be discharged by the owner to a safe environment!</p>

	<h2 style="margin: 0;">NOTE</h2>
	<ul style="list-style-type: none"> • Connection parts must be clean and free of residues. Contaminants may reach inside the combustion calorimeter and may cause false measurements and/or damage. • Calibration gas connection / compression fitting on side of combustion calorimeter • The input pressure for the gas connections must not exceed the specifications of the technical data of the combustion calorimeter. • Calibration gas must be free of contaminants and condensate. • The calibration gas quality must be similar to the process gas quality. • Each connection point must be carefully checked for leak tightness. When leaks are present, the system draws air and indicates false measured values. • Do not use sealant for sealing the gas connections. Sealant ingredients may falsify the measurement result. • Only suitable lines may be used. • For device versions with 2 or 3 measuring ranges, up to 2 calibration gases may be necessary. • Software-side configuration of the calibration gas. (☞ <i>Chapter 9.3 Available Displays</i>)

6.3.7 Flue gas

	<table border="1"><tr><td data-bbox="304 318 1361 427"> WARNING</td></tr><tr><td data-bbox="304 427 1361 949"><p>Risk of serious injury from escaping flue gases!</p><ul style="list-style-type: none">• Flue gases must be discharged to the outside!• For flue gases containing CO, H₂, and H₂S, adequate room ventilation must be ensured.</td></tr></table>	 WARNING	<p>Risk of serious injury from escaping flue gases!</p> <ul style="list-style-type: none">• Flue gases must be discharged to the outside!• For flue gases containing CO, H₂, and H₂S, adequate room ventilation must be ensured.
 WARNING			
<p>Risk of serious injury from escaping flue gases!</p> <ul style="list-style-type: none">• Flue gases must be discharged to the outside!• For flue gases containing CO, H₂, and H₂S, adequate room ventilation must be ensured.			

	<table border="1"><tr><td data-bbox="304 1016 1361 1126">NOTE</td></tr><tr><td data-bbox="304 1126 1361 1276"><p>Flue gases/residual heat must be discharged to the outside protected from air draft and without interruption through a vent.</p><p>An accumulation of residual heat leads to false measurement results.</p></td></tr></table>	NOTE	<p>Flue gases/residual heat must be discharged to the outside protected from air draft and without interruption through a vent.</p> <p>An accumulation of residual heat leads to false measurement results.</p>
NOTE			
<p>Flue gases/residual heat must be discharged to the outside protected from air draft and without interruption through a vent.</p> <p>An accumulation of residual heat leads to false measurement results.</p>			

6.3.8 Electrical connection

DANGER

Danger of electric shock!

Changes to the electrical equipment of the combustion calorimeter may be carried out only by skilled electricians in accordance with electrotechnical rules.

Parts of the open combustion calorimeter marked with the adjacent symbol may still carry voltage even when the main switch is switched off! If required, disconnect the combustion calorimeter from the supply network.







Fig. 6.3: Electrical connection plate, example

Connect the combustion calorimeter to the voltage supply using connections L1, N, PE in accordance with national requirements.

	 WARNING
	<p>After the line voltage is switched off, the device capacitors still carry voltage for up to 5 minutes! This time must be allowed to elapse before starting work on the high voltage electrical system.</p>

	NOTE
	<ul style="list-style-type: none">• Check whether the line voltage present is consistent with the device voltage of the power supply of the combustion calorimeter.• The combustion calorimeter may only be operated with effective protective earth connection in accordance with local requirements for high operating currents.

6.3.9 Electrical interfaces

	 WARNING
	<p>Endangerment of people and equipment when the combustion calorimeter is commissioned by non-instructed personnel! Allow only instructed/trained service technicians to carry out commissioning!</p>




Fig. 6.4: USB interface on the left side of the housing

The USB interface is used for transferring data. A memory stick is supplied with the device.

6.3.10 Connector assignment Input-Output IOexternal




Fig. 6.5: Connector assignment Input-Output IO type 06

	<h2 style="margin: 0;">WARNING</h2>
	<p>Do not connect to line voltage!</p> <p>Only operate outputs (relay, analog, digital) and inputs with safety extra-low voltage!</p> <p>Only connect serial interface – RS232 – to an electrically safe device!</p>

Relay outputs Connector X14

Digital output	Pin/Connector X14	Function	Status display
1	1 Common	Process	D 2
	2 Normally open		
	3 Normally closed		
2	4 Common	Maintenance	D 4
	5 Normally open		
	6 Normally closed		
3	7 Common	Filter change	D 5
	8 Normally open		
	9 Normally closed		
4	10 Common	Fault	D 6
	11 Normally open		
	12 Normally closed		
5	13 Common	Function, if applica- ble optionally assigned	D 7
	14 Normally open		
	15 Normally closed		
6	16 Common	Function, if applica- ble optionally assigned	D 8
	17 Normally open		
	18 Normally closed		
7	19 Common	Function, if applica- ble optionally assigned	D 9
	20 Normally open		
	21 Normally closed		
8	22 Common	Function, if applica- ble optionally assigned	D 10
	23 Normally open		
	24 Normally closed		

mA outputs Connector X5

Analog output	Pin/Connector X5	Function	Disconnecting module no./Jumper no. open
1	1 +mA	Wobbe	TR 1/JP 5
	2 -mA		
2	3 +mA	Density	TR 2/JP 6
	4 -mA		
3	5 +mA	Heating value	TR 3/JP 7
	6 -mA		
4	7 +mA	Function, if applica- ble optionally assigned	TR 4/JP 8
	8 -mA		
5	9 +mA	Function, if applica- ble optionally assigned	TR 5/JP 9
	10 -mA		
6	11 +mA	Function, if applica- ble optionally assigned	TR 6/JP 10
	12 -mA		
7	13 +mA	Function, if applica- ble optionally assigned	TR 7/JP 11
	14 -mA		
	15 n.c.		
	16 n.c.		

Digital control inputs Connector X3

Control inputs	Pin/Connector X3	Function	Status display diode
1	1	Start measurement	D 25
	2	Start measurement	
2	3	Start calibration	D 24
	4	Start calibration	
3	5	Function, if applicable	D 23
	6	optionally assigned	
4	7	Function, if applicable	D 22
	8	optionally assigned	
5	9	Function, if applicable	D 21
	10	optionally assigned	
6	11	Function, if applicable	D 20
	12	optionally assigned	
7	13	Function, if applicable	D 19
	14	optionally assigned	
8	15	Function, if applicable	D 18
	16	optionally assigned	


Serial interface RS232

Pin/Connector X11	Signal
	1 -mA analog input 1
	2 +mA analog input 1
	3 -mA analog input 2
	4 +mA analog input 2
	5 PT 100 air conditioner
	6 PT 100 air conditioner
	7 Not assigned
	8 RI
	9 RTS
	10 CTS
	11 DSR
	12 DTR
	13 TXD
	14 RXD
	15 DCD
	16 RS232 GND
Pin/Connector X1	Signal
	1 DCD
	2 RXD
	3 TXD
	4 DTR
	5 RS232 DND
	6 DSR
	7 RTS
	8 CT
	9 RI

6.4 Commissioning after installation

 A large yellow warning triangle with a black exclamation mark inside, set against a white background.	<p data-bbox="320 342 1345 421"> WARNING</p> <p data-bbox="320 443 1302 517">Endangerment of people and equipment when the combustion calorimeter is commissioned by non-instructed personnel!</p> <p data-bbox="320 526 1310 560">Allow only instructed/trained service technicians to carry out commissioning!</p>
---	--

6.4.1 Removing/attaching transport securing devices

	NOTE
<p>Before commissioning/transport of the combustion calorimeter, it must be ensured that all transport securing devices are removed/attached.</p>	

The following transport securing devices must be removed/attached within the combustion calorimeter:




Fig. 6.6: Transport securing devices

Item no.	Component	Type of securing device
1	Transport securing device for density measuring cell	4x cable ties; 1x hexagon socket screw 5x16
2	Transport securing device of pressure controller	2x retaining belt; 1x bubble wrap (inside)

Transport securing device in the pressure controller



Fig. 6.7: Transport securing devices of pressure controller

Transport securing device	Order of removal/attachment of transport securing device
Transport securing device of pressure controller	<ul style="list-style-type: none"> • Remove the retaining belt. • Open the protective flap (screw closure). • Remove the transport securing device (bubble wrap). • Reclose the protective flap.


Proceed in reverse order to reattach the transport securing device.

Transport securing device of density measuring cell



Transport securing device	Order of removal/attachment of transport securing device
Transport securing device of density measuring cell	<ul style="list-style-type: none"> • Remove the black cable ties from the center eyebolt. • Remove the white cable ties around the density measuring cell. • Loosen the eyebolt. • Density measuring cell must swing free.


Proceed in reverse order to reattach the transport securing device.


6.5 Documentation

	<h1 style="margin: 0;">NOTE</h1>
<p>UNION Instruments GmbH recommends keeping a maintenance manual and documenting all work and tests.</p>	

7 Commissioning/Switching on


	 <h3 style="margin: 0;">NOTICE</h3>
	<p>In order to establish start readiness, also establish the start readiness of linked system components according to their operating instructions!</p>

	<h3 style="margin: 0;">NOTE</h3>
	<p>For initial commissioning or before an extended operating shutdown, back up the device configuration.</p> <p>Have the backup made by service technicians or in response to a special service instruction.</p>


	<h3 style="margin: 0;">NOTE</h3>
	<p>The following table contains significantly shortened steps for commissioning after an extended downtime.</p> <p>To switch on the combustion calorimeter again after a short shutdown, some steps can be omitted: <i>☞ Right column!</i></p>

Steps	Com- mis- sioning	Switch- ing on
Check whether the transport securing device has been removed. The density cell must be able to swing freely at the springs.	X	
Check whether the transport securing device of the gas pressure controller has been removed. The foam material within the controller must be removed.	X	
Check whether environmental conditions (<i>☞ Technical data</i>) meet the requirements.	X	
Check whether the combustion calorimeter is securely mounted.	X	
Check whether the device is suitable for the process gas.	X	
Check whether the process gas is correct.	X	
Check whether the gas connections are correct and leak-tight.	X	
If applicable, check whether the calibration gas is correct.	X	
Furnish/switch on owner-side energy supplies and media supply.	X	

Commissioning/Switching on

Steps	Com- mis- sioning	Switch- ingon
Check whether the transport securing device has been removed. The density cell must be able to swing freely at the springs.	X	
Ensure voltage.	X	
Ensure that the door is closed (safety switch).	X	X
Switch on the main switch.	X	X
Establish start readiness of linked system components.	X	X
 When the combustion calorimeter has been switched off only temporarily, production can be resumed!		

8 Description of the work stations/operator control elements

	<h1>NOTE</h1>
<p>This chapter contains only elements for operation of the combustion calorimeter by the normal operator.</p>	




Fig. 8.1: Work stations

Item No.	Description	Function/Activity
1	Display	Displaying status and operation
2	Main switch	Switching the device On/Off



Description of the work stations/operator control elements

9 Operation

	<h1 style="margin: 0;">WARNING</h1>
	<p>Risk of injury!</p> <p>Only operate the combustion calorimeter when all lines are installed and have been checked for leak tightness in accordance with country-specific regulations.</p>

9.1 Operation of membrane keyboard/Description of display

The software controller is operated using a membrane keyboard. The buttons shown can be selected by pressing them.

	<h1 style="margin: 0;">NOTE</h1>
	<p>Damage to membrane keyboard!</p> <p>Operation with pointed or sharp objects may damage the membrane keyboard!</p>










Fig. 9.1: Operator control elements/Structure of display

Item No.	Description	Function
1	Numeric display	Output of current measured values
2	Display field	Information field
3	Arrows/Return	Arrow keys enable movement to an input field. The Return key confirms the entered value.
4.1	Menu	Menu key returns you to the main menu from any other menu level. The key simultaneously saves the input data in the memory.
4.2	Start	Start key starts the measurement.
4.3	Stop	Stop key stops the measurement without switching off the power supply.
4.4	ESC	ESC key cancels the current input operation regardless of the menu level.
5	Input	Input keys are used to input numeric data. A value field on the screen must be active for this.
5.1	Screenshot	This key creates and automatically saves a screenshot on the inserted memory stick.
6	Inspection window	Inspection window for monitoring the ignition or combustion process.
7	LED status display	Power: Device switched on Operation: only in the process or during calibration Service: Service needed (filter, temp., etc.) Fault: Major fault Flame: Flame is burning
8	Menu keys	Menu keys are described in the software. Their meaning changes according to which menu is selected. The function is labeled on the current screen. <i>☞ Chapter 9.2 Basic operation!</i>
9	Graphic display	Graphic display of current measured values
10	Curve values	Value display for selected curves.

9.2 Basic operation

The keys described in the following are used for operation of the combustion calorimeter on the part of the software.

Symbol	Function
	<p>Back:</p> <ul style="list-style-type: none"> Causes the menu to jump to the next higher menu level all the way back to the main menu.
	<p>Scroll:</p> <ul style="list-style-type: none"> Causes the display of other menus that cannot be displayed in the currently displayed screen due to limited space. The individual menus are displayed in a rolling manner over and over from the start.
	<p>Selection:</p> <ul style="list-style-type: none"> Enables a selection from a list.
	<p>Plus / Minus:</p> <ul style="list-style-type: none"> Causes the selected numbers/fields to be summed or unsummed.
	<p>Arrow:</p> <ul style="list-style-type: none"> Causes a jump to the next digit of a numeric input.

	NOTE
	<p>Other symbols not described above refer to the different menus. These are described in the corresponding screens.</p>

9.3 Available displays

The available displays and their function are described below. The displays are accessed using the menu and function keys shown in the chapter headings.

The structure shown in item 9.5 forms the basis of the controller. Different colors represent the different depth of the menu structure.

9.4 General information




Fig. 9.2: General information

Display (example values)	Information
V 4.39R09	Version number of the software
04.02.2015 10:05:09	Current date / time
Methane: 95.01 Vol%	Volume share of methane
mV=24.56	mV signal of the thermal battery
V=1.071	Voltage signal of the density measuring cell
T1=29.04/0.24	Inlet temperature into the thermal body / heating of the air
p=4.00	Air differential pressure
Fr=51.13	Controlled fan frequency
Stabi=0.011	Standard deviation of the measurement value over 2 minutes


NOTE

STD: During a calibration, 0.015 is typically achieved. The calibration is concluded after that.

9.5 Menu structure

9.6	Main menu	54
9.6.1	Main menu - Options	55
	Main menu - Options - I/O	56
	Main menu - Options - I/O - Analog outputs	57
	Main menu - Options - I/O - Digital outputs	58
	Main menu - Options - I/O - mA display	60
	Main menu - Options - I/O - Digital inputs	61
	Main menu - Options - I/O - Display	62
	Main menu - Options - Calibration	63
	Main menu - Options - Calibration - Configuration of calibration gas	64
	Main menu - Options - Calibration - Automatic calibration	65
	Main menu - Options - Calibration - Automatic calibration 2	66
	Main menu - Options - Calibration - Calibration	66
	Main menu - Options - Calibration - Save base calibration	66
	Main menu - Options - Calibration - Calibration limits	67
	Main menu - Options - System	68
	Main menu - Options - System - General	69
	Main menu - Options - System - Ignition	71
	Main menu - Options - System - Ignition	71
	Main menu - Options - System - Load factory settings	72
	Main menu - Options - System - CSV Export	72
	Main menu - Options - System - Hold signal On/Off	72
	Main menu - Options - System - Settings	73
	Main menu - Options - System - Settings - Date / Time	73
	Main menu - Options - System - Settings - Language	74
	Main menu - Options - System - Settings - Password	75
	Main menu - Options - System - Settings - Change color	76
	Main menu - Options - System - Settings - Hardware 1	77
	Main menu - Options - System - Service	79
9.6.2	Main menu - Trend	80
	Main menu - Graphic - Select time	81
	Main menu - Graphic - Select value	82
	Main menu - Graphic - Select signal	83
	Main menu - Graphic - Select unit	83
	Main menu - Graphic - Select curve	84
9.6.3	Main menu - Event list	85
9.6.4	Main menu - Device information	86

9.6 Main menu




The main menu is the standard display during active operation.

The following menus are accessed from the main menu:

- Options
- Trend
- Eventlist
- Device info

9.6.1 Main menu - Options



I/O

Configuration possibility for the following parameters:

- Analog outputs
- Digital (relay) outputs
- mA display
- Digital inputs
- Display

Calibration

Configuration possibility for the following parameters:

- Configuration of calibration gas
- Automatic calibration
- Automatic calibration 2
- Calibration
- Save base calibration
- Calibration limits

System


Configuration possibility for the following parameters:

- General
- Ignition
- Update
- Load factory settings
- CSV export
- Hold signal On/Off
- Settings (e.g., date, language, password)
- Disable system

Service

For service technician only

Main menu - Options - I/O



Options for configuration for the following parameters:

- Analog outputs
- Relay outputs
- Disp. mA
- Digital inputs
- Display

Main menu - Options - I/O - Analog outputs

W.I. i 12.840 kWh m ³	C.V. i 10.119 kWh m ³	V4.39R09 Process 04.02.2015 15:32:34
S.G. 0.6210 rel.	dp air : 4.00 mbar dp wobbe : 3.79 mbar int.press.: 11.90 mbar	T1 = 45.84 °C T2 = 61.97 °C T2-T1 = 16.12 °C

No	Signal		Unit	Range from	Range to
1	W.I. s	4 - 20	kWh/m ³	8.000	16.000
2	S.G.	4 - 20	kg/m ³	0.500	1.000
3	C.V. s	4 - 20	kWh/m ³	8.400	13.100
4	---				
5	---				
6	---				
7	---				

->Main ->Option ->I/O ->Configuration of analog outputs

The mA signals are configured in this menu.

Signal (list field)	Wobbe-index, density, heating value,
Type (list field)	4 – 20 mA
Unit (list field)	kcal/m ³ , relative
MB from MB to	value fields for numeric input

The following must be observed for this:

Units:

xxx/m³ and BTU/ft³ correspond to a gas temperature of 0 °C and a barometric pressure of 1013 mbar.

xxx/Sm³ corresponds to a gas temperature of 15 °C (60 °F) and a barometric pressure of 1013 mbar.

☞ xxx stands for MJ, kcal, or kWh.


BTU/ft³ corresponds to a gas temperature of 15 °C (60 °F) and a barometric pressure of 1013 mbar.

Main menu - Options - I/O - Digital outputs

W.I. i 12.868 kWh m ³	C.V. i 10.129 kWh m ³	V4.39R09 Process 04.02.2015 10:08:54
S.G. 0.6195 rel.	dp air : 4.00 mbar dp wobbe : 3.80 mbar int.press.: 11.94 mbar	dp air : 4.00 mbar dp wobbe : 3.80 mbar int.press.: 11.94 mbar

Configuration of relay outputs						
No	Signal	Opr tr	Unit	Value 1	Value 2	Zero pos.
1	Operation					low
2	Maintenance					low
3	Filter change					low
4	Fault					low
5	Operation					low
6	Operation delayed					low
7	---					
8	---					

* ->Main ->Option ->I/O ->Configuration of relay outputs



Digital signals (relay outputs, floating change-over contacts) are configured in this menu.

The following must be observed for this:

Zero position:

The zero position of the digital outputs can be freely selected: low/high.

Process:

The process gas solenoid valve is open and the flame is burning.

Calibration:

The calibration gas solenoid valve is open and the flame is burning.

Maintenance:

The device must be checked in a foreseeable time, but is still ready for operation. The reason for maintenance can be obtained from the "General information" field. Examples of reasons include:

- Filter change (air filter)
- No calibration gas (calibration canceled)
- Room temperature too high (air inlet temperature > 41 °C)
- Gas pressure too low
- Calibration deviation outside of tolerances
- The calibration was unstable and was canceled

Fault:

Device is no longer usable because

- Air differential pressure is too low (< 3.5 mbar)
- Sensor break (PT100, thermoelectric battery, temperature at burner is too high)
- Gas pressure is too low

Operation:

The flame is burning. Thermoelectric voltage of process or calibration gas is greater than the ignition threshold.

Operation delayed:

The flame is burning. Thermoelectric voltage of process or calibration gas is greater than the ignition threshold and the delay time has elapsed.

Overtemperature:

The flame is too hot. The Wobbe signal is greater than 76 mV, i.e., the temperature rise in the inner tube of the thermoelectric battery is greater than 50 °C.


Calibration deviation:

The value determined during base calibration is exceeded or fallen below.

Internal pressure:

Insufficient process or calibration gas.

Main menu - Options - I/O - mA display



The mA values for the currently active outputs are displayed in this menu. With the **Disp. mA** key, all channels can be displayed one after the other. The value can be obtained from the "General information" field.

Main menu - Options - I/O - Digital inputs

W.I. i 6.7866 kWh m ³	C.V. i 5.3208 kWh m ³	V4.39R09 Process 04.02.2015 10:10:06 Gas in 2 sec
S.G. 0.6146 rel.	dp air : 4.00 mbar dp wobbe : -0.17 mbar int.press.: 2.70 mbar	T1 = 45.84 °C T2 = 53.64 °C T2-T1 = 7.80 °C

Configuration of contact inputs		
No	Signal	Zero pos.
1	---	low
2	---	low
3	---	low

->Main ->Option ->I/O ->Configuration of contact inputs


Various digital signals are configured in this menu.
The following must be observed for this:

The pin assignment is preassigned.

☞ *Chapter 6.3.10 Connector assignment Input-Output IOexternal!*

Signals	Zero position	Contact	CWD2005
Start calibration	high	open	Calibration starts
		closed	No calibration
	low	open	No calibration
		closed	Calibration starts
Start measurement	high	open	Measuring starts
		closed	Measuring stops
	low	open	Measuring stops
		closed	Measuring starts
Hold signals	high	open	Hold mA
		closed	mA online
	low	open	mA online
		closed	Hold mA
Cancel calibration	high	open	Calibration cancellation
		closed	No calibration cancellation
	low	open	No calibration cancellation
		closed	Calibration cancellation

Main menu - Options - I/O - Display

W.I. i 7.0238 kWh/m ³	C.V. i 5.5289 kWh/m ³	V4.39R09 Process 04.02.2015 10:10:30																		
S.G. 0.6196 rel.	dp air : 4.00 mbar dp wobbe : 3.84 mbar int.press.: 11.94 mbar	T1 = 45.83 °C T2 = 55.44 °C T2-T1 = 9.62 °C																		
<table border="1"> <thead> <tr> <th colspan="3">Display configuration</th> </tr> <tr> <th>No</th> <th>Phys. value</th> <th>Unit</th> </tr> </thead> <tbody> <tr> <td>1</td> <td>W.I. i</td> <td>kWh/m³</td> </tr> <tr> <td>2</td> <td>S.G.</td> <td>Specific</td> </tr> <tr> <td>3</td> <td>C.V. i</td> <td>kWh/m³</td> </tr> <tr> <td>4</td> <td>Pressure</td> <td>mbar</td> </tr> </tbody> </table>			Display configuration			No	Phys. value	Unit	1	W.I. i	kWh/m ³	2	S.G.	Specific	3	C.V. i	kWh/m ³	4	Pressure	mbar
Display configuration																				
No	Phys. value	Unit																		
1	W.I. i	kWh/m ³																		
2	S.G.	Specific																		
3	C.V. i	kWh/m ³																		
4	Pressure	mbar																		
* ->Main ->Option ->I/O ->Display configuration																				
																				

The display of measured values is configured in this display.
A signal and a unit is assigned to each of the 4 display windows.

Main menu - Options - Calibration

W.I. i 11.954 kWh/m ³	C.V. i 9.4076 kWh/m ³	V4.39R09 Process 04.02.2015 10:10:59
S.G. 0.6192 rel.	dp air : 4.00 mbar dp wobbe : 3.81 mbar int.press.: 11.91 mbar	dp air : 4.00 mbar dp wobbe : 3.81 mbar int.press.: 11.91 mbar

W.I. i
kcal/m3
Dif y 150.000
Off y 11027.458
value 10278.876

C.V. s
kWh/m3
Dif y 0.120
Off y 11.341
value 10.450

T ein
Deg. C
Dif y 0.250
Off y 46.530
value 45.821

* ->Main ->Option ->Calib.

Config. calibr. gas	Automatic calibr.	Automatic calibr. 2	Calibrate
---------------------	-------------------	---------------------	-----------

W.I. i 12.100 kWh/m ³	C.V. i 9.5234 kWh/m ³	V4.39R09 Process 04.02.2015 10:11:04
S.G. 0.6194 rel.	dp air : 4.00 mbar dp wobbe : 3.81 mbar int.press.: 11.96 mbar	T1 = 45.82 °C T2 = 61.10 °C T2-T1 = 15.27 °C

W.I. i
kcal/m3
Dif y 150.000
Off y 11027.458
value 10404.030

C.V. s
kWh/m3
Dif y 0.120
Off y 11.341
value 10.579

T ein
Deg. C
Dif y 0.250
Off y 46.530
value 45.821

* ->Main ->Option ->Calib.

Store basic calibration	Calib. - limits
-------------------------	-----------------

Calibration values are configured in this menu.

Main menu - Options - Calibration - Configuration of calibration gas

Wobbe i 69.345 kcal m ³	Heizwert 68.918 kcal m ³	V4.08 Zuendung 06.09.2013 08:42:27
Dichte 0.9877 rel.	dp Luft : 4.00 mbar dp Wobbe : 0.01 mbar Int. Druck: 0.01 mbar	mV= 0.68 V=1.997 T1=29.05/ 0.38 p=4.00 Fr=53.75 Stab=0.260

Konfiguration Kalibriergase					
Nr	Einheit	Wobbe i	Wobbe s	Einheit	Dichte
1	kcal/m ³	2337	2546	relativ	0.737
2	BTU/ft ³	0.00	0.00	relativ	0.000

* ->Hauptmenue ->Optionen ->Calib. ->Konfiguration Kalibriergase

The calibration gas is configured in this menu.

The following must be observed for this:

The calibration gas is input as a Wobbe index (Wobbe i and Wobbe s) and as a relative density. These values are calculated from the gas components of the calibration gas.

The manufacturer always uses dry gas as the basis.

Units:

xxx/m³ and BTU/ft³ correspond to a gas temperature of 0 °C and a barometric pressure of 1013 mbar.

xxx/Sm³ corresponds to a gas temperature of 15 °C (60 °F) and a barometric pressure of 1013 mbar.

☞ xxx stands for MJ, kcal, or kWh.

BTU/ft³ corresponds to a gas temperature of 15 °C (60 °F) and a barometric pressure of 1013 mbar.

Main menu - Options - Calibration - Automatic calibration

W.I. i 12.624 kWh m ³	C.V. i 9.9370 kWh m ³	V4.39R09 Process 04.02.2015 10:11:45
S.G. 0.6195 rel.	dp air : 4.00 mbar dp wobbe : 3.80 mbar int.press.: 11.95 mbar	T1 = 45.82 °C T2 = 61.72 °C T2-T1 = 15.90 °C

Configuration auto calibration			
Program	Day	Time	Every
1	Every day	00:00	1
2	Every day	06:00	1
3	Every day	08:00	1
4	Every day	12:00	1
5	Every day	15:00	1
6	---	00:00	0
7	---	00:00	0
8	---	00:00	0
9	---	00:00	0
10	---	00:00	0

* ->Main ->Option ->Calib. ->Configuration auto calibration

The automatic calibration is configured in this menu.


The following must be observed for this:

Input:

Day is a list field (Su, Mo, Tu, We, etc.), Time and Cycle are value fields.

Duration of calibration:

Depending on the device type, the calibration duration is 10 - 20 min.


	NOTE
	<p>At a restart of the device or at a change of calibration gas, the calibration may have to be started multiple times until the pulse line is sufficiently flushed with calibration gas and the stability criterion is met.</p>

Main menu - Options - Calibration - Automatic calibration 2

W.I. i 12.681 kWh m ³	C.V. i 9.9808 kWh m ³	V4.39R09 Process 04.02.2015 10:12:02
S.G. 0.6194 rel.	dp air : 4.00 mbar dp wobbe : 3.82 mbar int.press.: 11.91 mbar	T1 = 45.82 °C T2 = 61.79 °C T2-T1 = 15.97 °C

Program 2	Situation	Value
1	---	0
2	---	0
3	---	0
4	---	0
5	---	0
6	---	0
7	---	0
8	---	0
9	---	0
10	---	0

* ->Main ->Option ->Calib. ->Configuration auto calibration 2



The automatic calibration that is defined by a situation is configured in this menu.

Criterion 1:

Automatic calibration after restart.

Criterion 2:

Automatic calibration at a defined change in ambient temperature compared to the last calibration.

Main menu - Options - Calibration - Calibration

The calibration is started with this key.
This stops the automatic calibrations.

Main menu - Options - Calibration - Save base calibration


The measured values are saved with this key. These values are reference values for additional calculations during the next calibration. Any deviations from these reference values are stored in the event list.

Main menu - Options - Calibration - Calibration limits

W.I. i 12.681 kWh m ³	C.V. i 9.9808 kWh m ³	V4.39R09 Process 04.02.2015 10:12:02																																				
S.G. 0.6194 rel.	dp air : 4.00 mbar dp wobbe : 3.82 mbar int.press.: 11.91 mbar	T1 = 45.82 °C T2 = 61.79 °C T2-T1 = 15.97 °C																																				
<table border="1"> <thead> <tr> <th colspan="3">Configuration auto calibration 2</th> </tr> <tr> <th>Program 2</th> <th>Situation</th> <th>Value</th> </tr> </thead> <tbody> <tr><td>1</td><td>---</td><td>0</td></tr> <tr><td>2</td><td>---</td><td>0</td></tr> <tr><td>3</td><td>---</td><td>0</td></tr> <tr><td>4</td><td>---</td><td>0</td></tr> <tr><td>5</td><td>---</td><td>0</td></tr> <tr><td>6</td><td>---</td><td>0</td></tr> <tr><td>7</td><td>---</td><td>0</td></tr> <tr><td>8</td><td>---</td><td>0</td></tr> <tr><td>9</td><td>---</td><td>0</td></tr> <tr><td>10</td><td>---</td><td>0</td></tr> </tbody> </table>			Configuration auto calibration 2			Program 2	Situation	Value	1	---	0	2	---	0	3	---	0	4	---	0	5	---	0	6	---	0	7	---	0	8	---	0	9	---	0	10	---	0
Configuration auto calibration 2																																						
Program 2	Situation	Value																																				
1	---	0																																				
2	---	0																																				
3	---	0																																				
4	---	0																																				
5	---	0																																				
6	---	0																																				
7	---	0																																				
8	---	0																																				
9	---	0																																				
10	---	0																																				
* ->Main ->Option ->Calib. ->Configuration auto calibration 2																																						

The calibration tolerances are set and the deviations from the base calibration are displayed in this menu. If the calibration values exceed the defined tolerances, this is indicated as a calibration deviation for digital outputs.

Main menu - Options - System



The basic configurations of the device, such as ignition, time of day, language, and code key, are specified in this menu.

Main menu - Options - System - General

The "Change signals after holding" command causes a smooth transition after calibration or removal of the signal holding state. A sudden rise or fall of the measured value in the analog output signal is avoided. The continuous transition is specified in seconds.

W.I. i 12.822 kWh m ³	C.V. i 10.092 kWh m ³	V4.39R09 Process 04.02.2015 10:14:41 Gas pressure
S.G. 0.6195 rel.	dp air : 4.00 mbar dp wobbe : 3.81 mbar int.press.: 11.92 mbar	dp air : 4.00 mbar dp wobbe : 3.81 mbar int.press.: 11.92 mbar

General settings	
Change signals after hold (in sec.)	120
Purge time after fan or instr. start-up	10
Time delay power down display	0
Display speed	150
Carrier gas Cal. cycles (in min)	0
delay operation	120
ADC calibration cycle	60
Calibration valve delay	0
Min. internal pressure (mbar)	8
Warning level internal pressure (mbar)	14

* ->Main ->Option ->System ->General settings

General system values are configured in this menu.

The following must be observed for this:

Change signals after holding:

When the Hold signals (mA) function is ended, an adjustment of the old and new measured values is made over a time ramp.

Purge time after cooling:

Specifies the length of time after the device start that the solenoid valve is open until ignition starts.

Screen switch-off time:

Specifies the time after which the screen will be switched off if no input is made.

Display speed:

Refers to a time constant for various burners. It is preset by the manufacturer and dependent on the burner type.

Carrier gas check cycles:

The time intervals for carrier gas calibrations, in hours, is defined.

Operation delay:

The "Operation" relay output is activated only after the delay time elapses.

Minimum internal pressure:

When the minimum internal pressure is fallen below, the device goes to STOP state, default value is 8 mbar.


Internal pressure warning threshold:

Below the warning threshold, insufficient gas is signaled for internal pressure and Service. Default value is 14 mbar.

Main menu - Options - System - Ignition

NOTE

The combustion calorimeter can only be ignited when the door is closed.



The ignition monitoring is configured in this menu.

The following must be observed for this:

Single ignition or interval ignition must be chosen.

Single ignition:

After device start and the flushing time period (10 s), the ignition starts for the maximum set ignition duration. This time can turn out to be shorter if the ignition threshold is reached before the ignition duration elapses. If the ignition threshold is not reached within the specified time, the combustion calorimeter goes to STOP state.

Interval ignition:

After device start and the flushing time period (10 sec), the ignition starts for a maximum set ignition duration and is repeated after a pause equal to the ignition duration. This is repeated until the ignition threshold is reached.

The ignition threshold specifies the differential temperature between the cooling air and flue gas. Default value is 3 °C.

Main menu - Options - System - Ignition

An update from a memory stick is initiated with this key.

Main menu - **Options** - **System** - **Load factory settings**

The factory settings are loaded with this key.

Main menu - **Options** - **System** - **CSV Export**

Three files are exported as an ASCII file with this key: 30 min at 1-second intervals, 3 hours at 10 second intervals, and 5 days at 1-minute intervals.

The values are separated by TAB. All possible 25 curves are written at once to a CSV file. This file can be processed, for example, with MS Excel.


Main menu - **Options** - **System** - **Hold signal On/Off**

The "Hold signal" function is switched on and off with this key.


Hold signal:

Storage of the mA values is activated with this function. After the function is ended, an adjustment of the old and new mA signals is made over a time ramp (120 sec).

Main menu - Options - System - Settings



Main menu - Options - System - Settings - Date / Time




The date and time are set in this menu.

- Increments selected numbers
- Decrements selected numbers
- Moves one field left
- Moves one field right


Main menu - Options - System - Settings - Language

The Language menu item contains 10 submenus for inserting, copying, and correcting various languages, even those that have to be represented by bitmaps (e.g., Chinese). Download languages onto a memory stick, correct them, and then re-import them to the combustion calorimeter.



The language is changed or your own bitmaps are configured in this menu.

Main menu - Options - System - Settings - Password



The password is changed in this menu.


The following must be observed for this:

Factory password:

The combustion calorimeter is delivered with factory password **0**. This can be changed, if necessary.

Unlock time:


This time specifies when the system will be locked and the password must be re-entered.

	<p>NOTE</p> <p>The time must be greater than 0. Otherwise, the system can then no longer be unlocked!</p>
---	--

Main menu - Options - System - Settings - Change color

W.I. i 12.864 kWh m ³	C.V. i 10.126 kWh m ³	V4.39R09 Process 04.02.2015 10:18:19
S.G. 0.6197 rel.	dp air : 4.00 mbar dp wobbe : 3.80 mbar int.press.: 11.94 mbar	mV=24.56 / V=1.071 Frequ. = 51.13 Hz Stabi. = 0.011
Color settings		
Menue Hintergrund	Black	
Menue Rahmen	Orange	
Menue Schrift	Grey 30	
Dialog Auswahl	White	
Kurve 1	Red	
Kurve 2	Green	
Kurve 3	Blue	
Kurve Auswahl	Yellow	

* ->Main ->Option ->System ->Settings ->Color settings



The display colors can be changed here.

Main menu - Options - System - Settings - Hardware 1

NOTE

Only change this menu in consultation with the manufacturer.

W.I. i <div style="font-size: 24pt; font-weight: bold;">12.869</div> <div style="font-size: 18pt;">kWh m³</div>	C.V. i <div style="font-size: 24pt; font-weight: bold; color: red;">10.130</div> <div style="font-size: 18pt; color: red;">kWh m³</div>	V4.39R09 Process 04.02.2015 10:19:41
S.G. <div style="font-size: 24pt; font-weight: bold;">0.6196</div> <div style="font-size: 18pt;">rel.</div>	dp air : 4.00 mbar dp wobbe : 3.81 mbar int.press.: 11.91 mbar	mV=24.56 / V=1.072 Freq. = 51.22 Hz Stabi. = 0.005

Configuration hardware 1				
Range No.	Dia.WJet	Range cfg.	Calibr. gas	Blending
1	0.00	0	1	180
2	0.00	0	1	Fan type
3	0.00	0	1	digital

No	Signal	Unit	Value	Switch (Z)	Bus type
1	---	kcal/Sm3	0.00	0.00	Union
2	---	kJ/m3	0.00	0.00	Bus Comm.
Change range				none	COM 2
Burner time constant				300	
Max. temp. inside tube (deg.C)				50.00	Def. RED
Calibration constant off				0.00	3

* ->Main ->Option ->System ->Settings ->Configuration hardware 1

The parameters for a multi-range measuring device are configured in this menu.

The screen shows all possible configurations.

Range No. indicates the number of installed measuring ranges. A maximum of 3 measuring ranges are possible. In this example, 3 measuring ranges are configured.

Dia. WJet. indicates the nozzle diameter in mm of the measuring range nozzle. This specification is information only and is meant to help later in calculating the gas volumes.

Range cfg. is a binary coding for the measurement conditions while a measuring range switchover is in process.

- 1 measuring range not loaded
- 1 nozzle set 2 or nozzle set 1
- 2 air nozzle on or off
- 4 flammable carrier gas on or off
- 8 not flammable carrier gas (oxygen) on or off
- 16 flammable carrier gas that does not burn by itself on or off

The number 6 indicates that the measuring range is operated with nozzle set 1 and a flammable carrier gas with an air nozzle.

The number 2 indicates that nozzle 1 has an air nozzle.

Calibr. gas indicates the number of the calibration gas for the measuring range. The upper nozzle set has number 1, the lower nozzle set the number 2.

No. indicates switchover point 1 and 2.

Signal indicates the physical unit of the measurement value.

Unit indicates the switchover point in the specified dimension between the measuring ranges. In the example, the first measuring range is switched over at 650 +5% kcal/m³.

Switch (%) indicates the overlap of the measuring ranges in %. In the example, the hysteresis is 5 %.

Measuring range switchover can be done automatically or manually. If manual measuring range switchover is set, an additional menu item appears under Number 6.2.1.3.9. Change range. Here you do the manual switchover to another measuring range.

- non
- automatic
- manual

The time constant of the burner is a factor determined for every burner. Sewage gas burners have another factor than burners for natural gas. Every burner type has its own coefficient.


Max temp of inner pipe is a switch-off temperature if in case of a fault highly calorific gas flows to a too large nozzle and generates a significant over-temperature.

Main menu - Options - Service




NOTE

Only change this menu in consultation with the manufacturer.



The menu is password-protected. Setting for Service only.

9.6.2 Main menu - Trend



The graphics are configured in this menu.


The following must be observed for this:
Up to three different curves in various colors can be displayed.

Selection:

A selection can be made from the following parameters:


- Select time
- Select value
- Select signal
- Select unit
- Select curve

Main menu - Graphic - Select time




The diagram can be optimally designed. The increments for increasing and reducing are programmed in order to obtain a resolution.

The diagram is subdivided into 10 segments on the x-axis. The segment time and the total displayed time can be set as follows:




Minimum segment time: 1 s
Maximum segment time: 12 h

The time values can be viewed as follows:



Main menu - Graphic - Select value




- Values for the curves can be set in this menu.

The curves can be viewed as follows:

- Ref y-value linearly to y-axis
- Ref y-value linearly from y-axis
- dy value compressed y-axis
- dy value stretched y-axis


Main menu - Graphic - Select signal



- Signals for the curves can be selected in this menu.
- Examples of signals:


Wobbe i Heating value mVSA Frequency T amb
Wobbe s Combustion mV p Wobbe T in
value

Main menu - Graphic - Select unit



- Units for the signals can be selected in this menu.

Main menu - Graphic - Select curve



Individual curves can be selected in this menu in order to change parameters.

9.6.3 Main menu - Event list

NOTE

The event list stores all events that are of interest for operation and service of the device.

A total of 1000 events can be stored.

W.I. i <div style="font-size: 24px; font-weight: bold;">12.874</div> <div style="font-size: 18px;">kWh m³</div>	C.V. i <div style="font-size: 24px; font-weight: bold; color: red;">10.138</div> <div style="font-size: 18px;">kWh m³</div>	V4.39R09 Process 04.02.2015 10:25:23
S.G. <div style="font-size: 24px; font-weight: bold;">0.6201</div> <div style="font-size: 18px;">rel.</div>	dp air : 4.00 mbar dp wobbe : 3.79 mbar int.press.: 11.90 mbar	mV=24.61 / V=1.073 Freq. = 51.02 Hz Stabi. = 0.001
<pre> 04.02.2015 10:10:23 Status operation 04.02.2015 10:10:21 Status ignition 04.02.2015 10:09:58 Status operation 04.02.2015 10:09:53 Status fault 04.02.2015 08:08:25 Status operation 04.02.2015 08:08:24 B. cali. deviat. (density) gas 1 04.02.2015 08:08:24 Calib. deviation (density) gas 1 04.02.2015 08:08:24 B. cali. deviation (wobbe) gas 1 04.02.2015 08:08:24 Calibr. deviation gas 1 04.02.2015 07:59:02 Calib. gas 1, AutoKal 1 04.02.2015 07:59:02 Status calibration gas 1 04.02.2015 06:08:32 Status operation 04.02.2015 06:08:31 B. cali. deviat. (density) gas 1 04.02.2015 06:08:31 Calib. deviation (density) gas 1 04.02.2015 06:08:31 B. cali. deviation (wobbe) gas 1 04.02.2015 06:08:31 Calibr. deviation gas 1 04.02.2015 05:59:10 Calib. gas 1, AutoKal 1 04.02.2015 05:59:10 Status calibration gas 1 04.02.2015 00:08:23 Status operation </pre>		
* ->Main		
<div style="display: flex; justify-content: space-between; align-items: center;"> ← <div style="display: flex; gap: 10px;"> <div style="border: 1px solid red; padding: 2px 5px; font-size: 8px;">Show start-up</div> <div style="border: 1px solid red; padding: 2px 5px; font-size: 8px;">Show calibration</div> <div style="border: 1px solid red; padding: 2px 5px; font-size: 8px;">Show all</div> </div> ↻ ↺ </div>		

- Show startup** Selection of the different events
- Show calibration** Selection of the different events
- Show all** Cancel selection

9.6.4 **Main menu** - **Device information**

NOTE

The values shown are very important for remote diagnostics when errors occur and can be loaded onto the memory stick and sent to the manufacturer by email, refer to CSV Export, Index.

CWD2005 - Device data

Type: CWD2005
 Gas type: -
 Carr.-gas: -
 Serial No.: -

Measurement ranges

Wobbe: -
 Density: 0,2 - 2,2

Configuration

Jet MR1:	Wobbe -	Jet MR2:	Wobbe -
	Air -		Air -
Burner:	-	Carr.-gas:	-

Additional options

- no additional options installed

Analog outputs

W.I. s	4	- 20mÅ	kWh/m3	8.000	-	15.000
S.G.	4	- 20mÅ	kg/m3	0.500	-	1.000
C.V. s	4	- 20mÅ	kWh/m3	8.400	-	13.100

 ->Main

Press any key to return to CWD2005 menu...

Distributor: TEST UI



UNION Instruments GmbH
 Zeppelinstr. 42
 76185 Karlsruhe
 GERMANY
 Tel.: +49-0721-95243-0
 Fax: +49-0721-95243-33
 Email: info@union-instruments.com



Last calibration deviation (max):
 04.02.2015 08:08:24
 -0.95 %


- 1 Device data:
 - Type
 - Gas type
 - Carrier gas
 - Device number
- 2 Measuring ranges:
 - Wobbe
 - Density
- 3 Equipment:
 - Nozzle MR1
 - Nozzle MR2
 - Burner
- 4 Additional options:
 - Installed additional options are listed here, if applicable.
- 5 Analog outputs:
 - Wobbe i
 - Density
 - Heating value




10 Decommissioning / Switching off

	 WARNING
	<p>Endangerment of people and equipment when the combustion calorimeter is decommissioned by non-instructed personnel!</p> <p>Allow only instructed/trained service technicians to carry out decommissioning!</p>

	 NOTICE
	<p>In order to decommission the combustion calorimeter, also decommission the linked system components according to their operating instructions!</p>

	NOTE
	<p>The following table contains steps for decommissioning for an extended downtime.</p> <p>To switch off the combustion calorimeter only temporarily, some steps can be omitted: <i>☞ Switching off column!</i></p>

Steps	Switching off	Decommissioning
Disconnect the device from the process, professionally close the line.	X	X
Bring linked system components to a standstill.	X	X
Switch off the main switch.	X	X
 If the combustion calorimeter is to be taken out of service only temporarily, the process ends here!		
Professionally disconnect/switch off owner-side energy supplies, media supply, and signal transmission.		X
When appropriate, pack the combustion calorimeter in a suitable manner. Ensure before starting or restarting transport that all transport securing devices are attached.		X










11 Maintenance

The quality of measurements by the combustion calorimeter can only be guaranteed when the maintenance intervals are adhered to.


11.1 Preparations

Supply lines of linked system components can be closed for maintenance purposes. These must be reopened after the device is put back into service.

	 DANGER
	<p>Risk of serious injury from electricity!</p> <ul style="list-style-type: none"> • Parts of the combustion calorimeter with the adjacent symbol may still carry voltage even when the main switch is switched off! If required, disconnect the combustion calorimeter from the supply network! • Switch off the main switch, and disconnect the device from the voltage supply if necessary, and take steps to prevent reconnection! • Only skilled electricians are permitted to work on the electrical equipment of the combustion calorimeter!

   	 WARNING
	<p>Risk of serious injury from escaping gases!</p> <ul style="list-style-type: none"> • Before carrying out maintenance work on the combustion calorimeter and whenever necessary, also bring the the linked system components to a standstill. • Gas connections may only be made by qualified personnel! Guidelines applicable at the installation location must be observed! • Incomplete combustion during maintenance operation may cause the exhaust air to become polluted with process gas! • In the case of toxic gases, the applicable safety provisions must be complied with. • Risk of serious injury from burns caused by burner components! Before carrying out maintenance work on the burner system, always wait 15 minutes for it to cool down.

11.2 Maintenance work/Inspection

	NOTE
	Maintenance work must be performed according to the inspection and maintenance schedule! The nature and amount of wear depends greatly on the individual use and operating conditions. All specified intervals are therefore guide values.

The following items must be ensured before carrying out maintenance work:

1. Make a note of the following values on the combustion calorimeter!
 - Wobbe i/s
 - Heating value/combustion value
 - Density
 - mV signals
 - mA signals
 - Internal pressure
 - Differential pressure, Wobbe
 - Differential pressure, air
 - Frequency of the fan control
2. Notify the control room
3. If no default values via PLC are possible, activate "Hold signal" on the combustion calorimeter.
4. Close the shut-off valve on the combustion calorimeter after a leak test.

The nature and amount of wear depends greatly on the individual use and operating conditions. All specified intervals are therefore guide values.

To guarantee operational reliability, use only genuine spare parts of the manufacturer.

Check	Interval (recommended)
-------	------------------------

After commissioning

Check and, if necessary, update the firmware version.	As required
Store the current configuration	As required


Semi-annual check

Check air filter for contamination/clogging.	Every 6 months
Perform calibration (at shorter intervals depending on accuracy requirements)	As required
Check tube connection to the density measuring cell/sensors for porosity.	Every 6 months
Check membrane on the gas pressure controller for porosity or lack of elasticity.	Every 6 months
Remove ² /clean ³ thermoelectric battery and heat exchanger.	Every 6 months
Replace the seal on the heat exchanger.	Every 6 months
Check the fan.	Every 6 months
Check the air intake filter (ambient air inlet, fan filter mat).	Every 6 months
Clean ⁴ the gas and air nozzles Change O-rings, if necessary.	Every 6 months

Annual check

Replace air filter.	Annually
Replace neoprene tubes inside the combustion calorimeter.	Annually
Clean the complete system.	Annually

Additional information:

 *Included documents*





² Flow plates are attached on the side of the thermoelectric battery and in the thermal elements of the combustion calorimeter. The flow plates must not be twisted when removing the thermoelectric battery. The thermoelectric battery must be pulled out as far as possible and then tilted to the side.


³ Cleaning of the heat exchanger with water followed by thorough drying of the cleaned components.

⁴ Clean the nozzles using a volatile solvent.



12 Troubleshooting

   	<div style="background-color: #f4a460; padding: 5px;">WARNING</div> <p>Risk of serious injury from electricity and escaping gases!</p> <ul style="list-style-type: none"> • Before carrying out maintenance work on the combustion calorimeter and whenever necessary, also bring the the linked system components to a standstill. • Switch off the main switch, and disconnect the device from the voltage supply if necessary, and take steps to prevent the main switch from being switched on again and the voltage supply from being reconnected! • Only skilled electricians are permitted to work on the electrical equipment of the combustion calorimeter! • Parts of the combustion calorimeter with the adjacent symbol may still carry voltage even when the main switch is switched off! • If required, disconnect the combustion calorimeter from the supply network!
--	---

	<div style="background-color: #0070c0; color: white; padding: 5px;">NOTE</div> <p>Troubleshooting is broken down into the following categories:</p> <ul style="list-style-type: none"> • Unstable measured value • Drift of measured value • Incorrect ignition behavior
---	--

12.1 Preparations for troubleshooting

Supply lines of linked system components can be closed for maintenance purposes. These must be reopened after the device is put back into service.

NOTE



Event list:

The software keeps an event list with up to 1000 events in chronological order (specification of date). The event list provides information about incorrect behavior, refer to 9.6.3.

The export of event data to a special data carrier (memory stick) is possible after consultation with Service. The exported data can be sent to the manufacturer for fault analysis, refer to CSV Export, 9.6.1.

Events are: Start, Stop, Ignition, Insufficient cooling air, Calibration

12.2 Changing/replacing fuses

Only skilled electricians or service technicians are permitted to replace fuses. Only replace with fuse types specified by UNION.

12.3 Unstable measured value

- The inlet pressure controller cannot maintain a constant inlet pressure. The process pressure (blast furnace gas) is too high. A pressure booster pump must be installed.
- Direct solar radiation is causing rapid temperature changes. Direct solar radiation must be prevented.
- Air condition is causing the temperature to change too quickly. The cooling capacity of the air conditioning is too high and the hysteresis is too large.
- The inlet pressure is inexact/cannot be maintained. Check whether the input pressure is too high.

12.4 Drift of measured value

The measured value drifts upward in one direction:

- The calibration no longer achieves the required point. The frequency controller is at its maximum value.
- Heavy filter contamination that can no longer be corrected. The filter must be replaced.

The measured value drifts downward:

- The heat exchanger is contaminated with soot (when propane or butane is combusted with too little air)
- The heat exchanger must be cleaned (hot water) and then dried with compressed air.
- The thermal battery is contaminated.
- The thermoelectric battery must be cleaned and carefully dried.

12.4.1 Incorrect ignition

The combustion calorimeter is continually igniting, the flame is burning. Combustion calorimeter does not switch to the operation state.

- Incorrect temperature setting (too high). The temperature threshold must be reduced.

The combustion calorimeter switches to the operation state even though the flame is not burning and then reverts to the ignition state.

- Incorrect temperature setting (too low). The temperature threshold must be increased.

Ignition electrodes are corroded, natural wear when ignition is frequent.

- Replace the ignition electrodes.

12.4.2 Error/status messages

Filter change

- Filter change (air filter)
- The frequency of the frequency converter is > than the set threshold value.

Int. pressure

- Alarm in the event of inadequate process gas or calibration gas
- Adjustable threshold; default: 14 mbar

Overtemp

- The flame is too hot. Wobbe signal > 76 mV,
- Temperature rise in the inner tube of the thermoelectric battery > 50 °C

Fault

- Filter change of air filter (the frequency of the frequency converter is > than the set threshold value)
- Gas pressure is too low
- Fault in the thermoelectric battery
- Defective PT 100 temperature sensor

Service

Filter change of air filter

- Frequency > threshold value
- Air inlet temperature > 41 °C

Operation

- Flame is burning, analyzer is running on process or calibration gas

Operation delay

- Flame is burning, analyzer is running on process or calibration gas
- Delay time has elapsed after which the analog signals are enabled

Process


- Flame is burning, solenoid valve for process gas is open
- Analyzer is running on process gas

Calibration

- Flame is burning, solenoid valve for calibration gas is open
- Calibration is active, no mA signals are output, these are "on hold", the last current value is retained.



13 Service

NOTE	
	<p>UNION Instruments GmbH is available to answer any questions.</p> <p>In case of orders or technical questions, please provide your customer number, phone number where you can be reached, the combustion calorimeter type and number (see name plate), and required spare parts/bills of material numbers, if applicable.</p>

UNION Instruments GmbH - Service

Maria-Goeppert-Straße 22
23562 Lübeck
Germany



+49 (0)721-680381-30



+49 (0)721-680381-33



support@union-instruments.com



<http://www.union-instruments.com>




14 Disposal



In case of decommissioning, a return of the device to UNION Instruments GmbH is possible.

Suggestion: Have UNION Instruments GmbH dispose of your combustion calorimeter.

	 WARNING
	Risk of injury from electricity and, if applicable, gases in the combustion calorimeter!

 Umweltgefährlich	NOTE
	Observe national regulations for disposal of machines and working materials. Sort parts by groups and supply them to professional recyclers.

15 Spare parts

	 WARNING
	<p>Use of non-approved spare parts (e.g., parts of other manufacturers, parts with deviating specifications, imitation consumables and wear parts) may cause damage and endanger people! Any warranty is voided in this case. The owner is then liable for damage that occurs!</p> <p>When standard components are replaced, use only identical components of the original manufacturer! In the event that components are discontinued or components of other manufacturers are used, this requires manufacturer's approval by UNION Instruments GmbH!</p>

Spare parts can be ordered from UNION Instruments GmbH:
☞ *Chapter 11 Service.*

Make a note of the combustion calorimeter type and number (☞ *Name plate*).
Identify and make a note of the order number, if applicable (☞ *Included documents*).

Order part.

The following spare part packages are among those available:

Spare part package 1 year operation
Spare part package 2-3 year operation

16 Appendix

Data Structure Profibus

Nr.	Typ	Bytes	Bytes tot.	Name	Description	Source
1	FLOAT	4	4	Heating value	measuring unit: kJ/m ³	CWD ⁵
2	FLOAT	4	8	calorific value	measuring unit: kJ/m ³	CWD
3	FLOAT	4	12	Wobbe i	lower Wobbe index: measuring unit:: kJ/m ³	CWD
4	FLOAT	4	16	Wobbe s	upper Wobbe index: measuring unit:: kJ/m ³	CWD
5	FLOAT	4	20	Dv	relativ density, without dimension	CWD
6	FLOAT	4	24	Reserve		CWD
7	FLOAT	4	28	Reserve		CWD
8	FLOAT	4	32	Tein	Air temp. CWD thermo, unit °C	CWD
9	FLOAT	4	36	CO	Unit: % (only CWDs with CO modul)	CWD
10	FLOAT	4	40	Air min	Min. Luftbedarf, Einheit: m ³ Luft / m ³ Gas (nur bei Kundenspezifischen CWDs)	CWD
11	FLOAT	4	76	Reserve		CWD
...						
20	FLOAT	4	80	Fixed test value	send: 12345,6789	CWD
21	BYTE	1	81	Operating condition	Operating condition CWD: 0 – Stop 1 – Proces 2 – Ignition 3 – Ignition pause 4 – Error 5 – over temperature 6 – calibration measuring range 1 7 – calibration measuring range 2 8 – calibration measuring range 3 9 – Test-Gas measurement (only in process)	CWD
22	BYTE	1	82	Zustand Filterwechsel	1 = Filterwechsel am CWD benötigt	CWD
23	BYTE	1	92	Reserve		CWD

Table 1: Data structure Profibus

For more information about communication with Profibus and the UNION calorimeter see Service UNION Instruments GmbH, *☞ Kapitel 11 Service.*

⁵ CWD = Calorimeter

Data structure serial interface RS 232 (option)

With the optional serial interface (RS 232) data can be transmitted. Parameters are selected in the "I / O Analog outputs". The interface has to be activated in the menu service.

Data transmission:	Transmission rate:	9600 baud
Parity-bit:	no	
Stop-bit:	1	
Data-bit:	8	

Output format

111111.111 222222.222 333333.333 777777.777

111111.111	=	Output data 1
222222.222	=	Output data 2
... ..		
777777.777	=	Output data 7

Assignment of the plugs/oins, ↗ 6.3.10

Index

A		M	
Accessories	26	Main switch	19
C		Maintenance.....	91
Commissioning	45	Maintenance work	92
Connecting	29	O	
Connections.....	23	Operation	49
Contact		Operator control elements	47
Service	101	P	
UNION Instruments GmbH.....	17	Personnel and qualification	14
CSV Export.....	72	Profibus	
D		Data structure	107
Decommissioning	89, 103	Protective equipment.....	19
Displays	51	S	
Disposal.....	103	Safety information	13, 15
E		serial interface	
Electrical connection	37	data struture	108
Electrical interfaces.....	38	Service	101
Environmental conditions	28	Spare parts.....	105
F		Symbols	13
Fault correction	95	T	
I		Transport.....	27
Inspection	92	Troubleshooting.....	95
Installation location	29	W	
Installing	29	Wall mounting	31
Intended use	14	Warning information	13

List of figures

Fig. 1.1:	Housing dimensions	5
Fig. 1.2:	Markings and warning information.....	6
Fig. 4.1:	Main switch (example).....	19
Fig. 4.2:	Figure Safety switch (example)	19
Fig. 4.3:	Markings and warning information (example)	21
Fig. 4.4:	Warning information on electrical connection plate (example)	21
Fig. 5.1:	Housing connections	23
Fig. 5.2:	Housing without doors, example	25
Fig. 6.1:	Wall mounting	32
Fig. 6.2:	Dimensions wall mounting.....	32
Fig. 6.3:	Electrical connection plate, example	37
Fig. 6.4:	USB interface on the left side of the housing.....	38
Fig. 6.5:	Connector assignment Input-Output IO type 06.....	39
Fig. 6.6:	Transport securing devices	43
Fig. 6.7:	Transport securing devices of pressure controller	44
Fig. 8.1:	Work stations	47
Fig. 9.1:	Operator control elements/Structure of display.....	49
Fig. 9.2:	General information.....	52

List of tables

Table 1:	Data structure Profibus.....	107
----------	------------------------------	-----